




Attachment no. 1 to the certificate of analysis for work order PR20B1201

Results of soil texture analysis



Test method specification: CZ_SOP_D06_07_120 Grain size analysis using the wet sieve analysis using laser diffraction (fraction from 2 µm to 63 mm) Fraction > 0.063 mm determined by wet sieving method, other fractions determined from the fraction "< 0.063mm" by laser particle size analyzer using liquid dispersion mode. Fractions "Sand >63 µm", "Silt 2-63 µm" and "Clay <2 µm" evaluated from measured data.

The end of result part of the attachment the certificate of analysis